## Supplement

Title:	56x Getting Starte	ed	
PN:	3833028	Supplement Iss	sue: 2
Print Date:	Dec. 2010	Issue Date:	2/13
Revision/D	ate:	Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

© 2011-2013 Fluke Corporation. All rights reserved.



## Change #1, 58989, 64912

On page 10, replace the *Specification Summary* table with:

## **Specifications Summary**

See Users Manual on CD for full specifications.

Feature	561	566	568
IR Temperature Range	-40 °C to 550 °C (-40 °F to 1022 °F)	-40 °C to 650 °C (-40 °F to 1202 °F)	-40 °C to 800 °C (-40 °F to 1472 °F)
Accuracy	$ \begin{array}{l} <0 \ ^\circ C: \pm (1 \ ^\circ C + 0.1 \ ^\circ /1 \ ^\circ C); \\ \geq 0 \ ^\circ C: \pm 1 \ \% \ of \ reading \ or \ \pm 1 \ ^\circ C, \ whichever \ is \\ greater \\ <32 \ ^\circ F: \pm (2 \ ^\circ F \ + 0.1 \ ^\circ /1 \ ^\circ F); \\ \geq 32 \ ^\circ F: \pm 1 \ \% \ of \ reading \ or \ \pm 2 \ ^\circ F, \ whichever \ is \\ greater \\ \end{array} $		
K-Type Thermo- couple Input Temperature Range	-40 °C to 550 °C (-40 °F to 1022 °F)	-270 °C to 1372 °C (-454 °F to 2501 °F)	
K-Type Thermo- couple Input Temperature Accuracy	Input accuracy $\pm 1$ % of reading or $\pm 2$ °C ( $\pm 1$ % of reading or $\pm 4$ °F), whichever is greater	$\begin{array}{c} -270 \ ^\circ C \ to \ -40 \ ^\circ C: \ \pm(1 \ ^\circ C \ + \\ 0.2 \ ^\circ/1 \ ^\circ C) \\ (-454 \ ^\circ F \ to \ -40 \ ^\circ F: \ \pm(2 \ ^\circ F \ + \ 0.2 \ ^\circ/1 \ ^\circ F)) \\ -40 \ ^\circ C \ to \ 1372 \ ^\circ C: \ \pm 1 \ \% \ of \\ reading \ or \ \pm 1 \ ^\circ C, \ whichever \ is \\ greater \\ (-40 \ ^\circ F \ to \ 2501 \ ^\circ F: \ \pm 1 \ \% \ of \\ reading \ or \ \pm 2 \ ^\circ F, \ whichever \ is \\ greater) \end{array}$	