

Supplement

Title: 700PEX I.S.	Supplement Issue: 4
Part Number: 2106534	Issue Date: 8/10
Print Date: May 2004	Page Count: 2
Revision/Date:	

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

Change #1

Under **Product Compliance Markings**, change the following:

Replace:

CE  II 1 G Eex ia IIC T4
0344

With:

CE  II 1 G EEx ia IIC T4
0344 KEMA 04ATEX1102 X

Replace:

Ex Certification by Mensor Corporation,
San Marcos, TX USA

With:

Manufactured by Mensor Corporation, 201
Barnes Drive, San Marcos, TX USA 78666

Change #2

Under **Product Compliance Markings**, add the following:



Conforms to Australian directives.

Change #3

Under **Safety Information**, add the following to the **Warning** section:

- Entity parameters for intrinsic safety can be found in Table 3.

Under **Faults and Damage**, add the following to the **Warning** section:

- Substitution of components may impair intrinsic safety.

Change #4

Replace the **Box Contents** with:

Box Contents

Each pressure module comes with an installed strap and instruction sheet. All pressure modules, with the exception of the 700P29Ex, come, with ¼ NPT F to ISO7-R ¼ (Metric) M adapters.