

Certificate of Calibration



Issued by: Megger Instruments Limited
Archcliffe Road
Dover
Kent CT17 9EN
www.megger.com

Instrument Details

Model DLRO100EB
Description Megger Digital Low Resistance Ohmmeter
Serial Number 1003101101468693

Environment Details

Temperature	Relative Humidity	Mains Frequency	Mains Voltage	Date Tested	22-Dec-2015
20°C ± 3°C	50% ± 20%	50Hz ± 0.5Hz	230V -2% + 10%	Date Commissioned	<input type="text"/>

Traceability Information

Serial No	Test Equipment Description	Procedure Version
T3068	DLRO100A AutoCal Test Equipment	V1.02

Test Title	Applied Value	Reading Value	Tolerance	Uncertainties	Pass/Fail
Resistance Ranges					
150µΩ @ 100A	151.83 µΩ	151.8 µΩ	0.5µΩ	±0.42µΩ	Pass
1mΩ @ 10A	1.0191 mΩ	1.0193 mΩ	3µΩ	±0.08µΩ	Pass
1mΩ @ 50A	1.0191 mΩ	1.0188 mΩ	4µΩ	±0.08µΩ	Pass
1mΩ @ 100A	1.0191 mΩ	1.0187 mΩ	4µΩ	±0.08µΩ	Pass

SAMPLE