

CERTIFICATION OF FUNCTION TEST & FACTORY CALIBRATION

Final test passed-

ROTRONIC AG

Device Type:	CP11
Serial No:	
Reference System:	HC2-S (SCS CERTIFIED)

ROTRONIC AG certifies that this instrument meets the specifications of our datasheet. It has been calibrated and corresponds to the test requirements of ISO9001-2008, the reference and service norms are traceable to national standards. The calibrated values are valid under above mentioned conditions only at the time of measurement and are referenced to the indicated references and working standards.

Mode	Production
Display	OK
Backlight	OK
Keys	OK
Power 5VDC	OK
Communication	OK
Humidity & Temp.	19.3%rh / 25.1°C 93.6%rh / 24.8°C (CP11 device)
	Reference : 19.4%rh / 25.2°C 94.2%rh / 24.7°C (HC2-S SCS certified)
CO2	528 ppm (CP11 device) Reference : 533 ppm (400ppm standard gas certified)