

## **Notes**:

- Exposure to direct sunlight can cause overheating and moisture condensation which will reduce the operating life of electronics.
- Protect Instruments from direct sunlight with this iridite finished aluminum sun screen (Greyline Option SCR).
- Seal conduit entries with caulking compound to further reduce moisture condensation.



INFO@PULSARMEASUREMENT.COM

Pulsar Measurement is a trading name of Pulsar Process Measurement, Ltd.

Copyright © 2021 Pulsar Measurement Registered Address: 1 Chamberlain Square CS, Birmingham B3 3AX Registered No.: 3345604 England & Wales

## Delivering the Measure of Possibility

**United States** 

+1 888-473-9546

A cia

+60 102 591 332

Canada

+1 855-300-9151

Oceania

+61 428 692 274

**United Kingdom** +44 (0) 1684 891371

pulsarmeasurement.com