

We,

GE Sensing
1100 Technology Park Drive
Billerica, MA 01821
USA

declare under our sole responsibility that the

OxyTraK™ 390 Flue Gas Oxygen Analyzer
OxyTraK™ 411 In Situ Flue Gas Oxygen Analyzer

to which this declaration relates, are in conformity with the following standards:

- EN 61326:1998, Class A, Annex A, Continuous Unmonitored Operation
- EN61010-1:1993 + A2:0995, Overvoltage Category II, Pollution Degree 2

following the provisions of the

- Electromagnetic Compatibility Directive 89/336/EEC
- Low Voltage Directive 73/23/EEC

April 23, 2007

Date of Issue



Mr. Gary Kozinski
Certification & Standards, Lead Engineer

